

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031939.D
 Acq On : 06 Oct 2022 23:34
 Operator : JC/MD
 Sample : N4883-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-135

Quant Time: Oct 07 08:09:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	51882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	86000	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	93533	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56951	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58538	54.633	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.260%	
35) Dibromofluoromethane	5.391	113	48958	50.292	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.580%	
50) Toluene-d8	8.653	98	178173	50.865	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.720%	
62) 4-Bromofluorobenzene	11.079	95	60961	49.227	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.460%	
Target Compounds						Qvalue
16) Acetone	2.386	43	2971	8.159	ug/l	95
65) Chlorobenzene	10.085	112	1511	0.533	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031939.D
Acq On : 06 Oct 2022 23:34
Operator : JC/MD
Sample : N4883-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-135

Quant Time: Oct 07 08:09:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

