

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031940.D
 Acq On : 06 Oct 2022 23:57
 Operator : JC/MD
 Sample : N4883-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-136

Quant Time: Oct 07 08:10:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	56205	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	93030	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62992	54.268	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.540%
35) Dibromofluoromethane	5.385	113	51815	49.205	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.400%
50) Toluene-d8	8.647	98	187563	49.499	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.000%
62) 4-Bromofluorobenzene	11.079	95	64306	48.004	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.000%
Target Compounds						Qvalue
11) Tert butyl alcohol	2.977	59	176814	1224.328	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	2052996	714.612	ug/l	99
22) Diisopropyl ether	3.758	45	3335	1.218	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031940.D
Acq On : 06 Oct 2022 23:57
Operator : JC/MD
Sample : N4883-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-136

Quant Time: Oct 07 08:10:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

