

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031941.D
 Acq On : 07 Oct 2022 00:20
 Operator : JC/MD
 Sample : N4883-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-137

Quant Time: Oct 07 08:10:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	55780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	93508	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60104	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63304	54.953	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.900%
35) Dibromofluoromethane	5.385	113	52441	49.545	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.080%
50) Toluene-d8	8.647	98	188329	49.447	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.900%
62) 4-Bromofluorobenzene	11.079	95	64108	47.612	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.220%
Target Compounds						
19) Methyl tert-butyl Ether	3.111	73	99812	35.008	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031941.D
Acq On : 07 Oct 2022 00:20
Operator : JC/MD
Sample : N4883-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-137

Quant Time: Oct 07 08:10:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

