

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031944.D
 Acq On : 07 Oct 2022 01:29
 Operator : JC/MD
 Sample : N4883-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-140

Quant Time: Oct 07 08:10:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	52895	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	88438	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	94552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57572	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60197	55.106	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.220%
35) Dibromofluoromethane	5.385	113	50629	50.575	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.140%
50) Toluene-d8	8.653	98	181447	50.371	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.740%
62) 4-Bromofluorobenzene	11.079	95	61744	48.485	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.960%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1319	3.553	ug/l #	85
19) Methyl tert-butyl Ether	3.111	73	4722	1.747	ug/l #	90
20) Methylene Chloride	2.782	84	941	0.965	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031944.D
Acq On : 07 Oct 2022 01:29
Operator : JC/MD
Sample : N4883-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-140

Quant Time: Oct 07 08:10:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

