

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031920.D
 Acq On : 06 Oct 2022 15:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 07 08:06:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	50.000	48.768	2.5	88	0.00
3 P	Chloromethane	50.000	45.934	8.1	91	0.00
4 C	Vinyl Chloride	50.000	45.016	10.0#	87	0.00
5 T	Bromomethane	50.000	54.450	-8.9	99	0.00
6 T	Chloroethane	50.000	51.137	-2.3	90	0.00
7 T	Trichlorofluoromethane	50.000	45.245	9.5	88	0.00
8 T	Diethyl Ether	50.000	50.834	-1.7	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.697	4.6	91	0.00
10 T	Methyl Iodide	50.000	42.570	14.9	78	0.00
11 T	Tert butyl alcohol	250.000	273.106	-9.2	108	-0.01
12 CM	1,1-Dichloroethene	50.000	46.341	7.3#	88	0.00
13 T	Acrolein	250.000	257.881	-3.2	93	0.00
14 T	Allyl chloride	50.000	48.691	2.6	90	0.00
15 T	Acrylonitrile	250.000	252.907	-1.2	96	0.00
16 T	Acetone	250.000	234.224	6.3	96	0.00
17 T	Carbon Disulfide	50.000	45.794	8.4	86	0.00
18 T	Methyl Acetate	50.000	50.932	-1.9	97	0.00
19 T	Methyl tert-butyl Ether	50.000	52.175	-4.3	96	0.00
20 T	Methylene Chloride	50.000	45.136	9.7	92	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.006	4.0	91	0.00
22 T	Diisopropyl ether	50.000	51.087	-2.2	95	0.00
23 T	Vinyl Acetate	250.000	271.887	-8.8	96	0.00
24 P	1,1-Dichloroethane	50.000	47.948	4.1	91	0.00
25 T	2-Butanone	250.000	250.165	-0.1	96	0.00
26 T	2,2-Dichloropropane	50.000	48.101	3.8	86	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.818	2.4	92	0.00
28 T	Bromochloromethane	50.000	49.760	0.5	92	0.00
29 T	Tetrahydrofuran	250.000	249.638	0.1	95	0.00
30 C	Chloroform	50.000	49.259	1.5#	92	0.00
31 T	Cyclohexane	50.000	47.510	5.0	89	0.00
32 T	1,1,1-Trichloroethane	50.000	48.690	2.6	88	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.583	2.8	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	49.292	1.4	94	0.00
36 T	1,1-Dichloropropene	50.000	48.620	2.8	90	0.00
37 T	Ethyl Acetate	50.000	51.968	-3.9	96	0.00
38 T	Carbon Tetrachloride	50.000	50.221	-0.4	87	0.00
39 T	Methylcyclohexane	50.000	50.215	-0.4	90	0.00
40 TM	Benzene	50.000	49.254	1.5	91	0.00
41 T	Methacrylonitrile	50.000	53.853	-7.7	97	0.00
42 TM	1,2-Dichloroethane	50.000	51.562	-3.1	95	0.00
43 T	Isopropyl Acetate	50.000	54.776	-9.6	97	0.00
44 TM	Trichloroethene	50.000	50.001	-0.0	91	0.00
45 C	1,2-Dichloropropane	50.000	51.512	-3.0#	94	0.00
46 T	Dibromomethane	50.000	52.590	-5.2	96	0.00
47 T	Bromodichloromethane	50.000	55.169	-10.3	94	0.00
48 T	Methyl methacrylate	50.000	54.834	-9.7	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031920.D
 Acq On : 06 Oct 2022 15:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 07 08:06:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1108.355	-10.8	94	0.00
50 S	Toluene-d8	50.000	49.267	1.5	93	0.00
51 T	4-Methyl-2-Pentanone	250.000	272.208	-8.9	97	0.00
52 CM	Toluene	50.000	51.194	-2.4#	91	0.00
53 T	t-1,3-Dichloropropene	50.000	51.793	-3.6	93	0.00
54 T	cis-1,3-Dichloropropene	50.000	57.566	-15.1	93	0.00
55 T	1,1,2-Trichloroethane	50.000	51.378	-2.8	95	0.00
56 T	Ethyl methacrylate	50.000	57.313	-14.6	96	0.00
57 T	1,3-Dichloropropane	50.000	51.965	-3.9	96	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	278.066	-11.2	98	0.00
59 T	2-Hexanone	250.000	271.106	-8.4	96	0.00
60 T	Dibromochloromethane	50.000	59.038	-18.1	95	0.00
61 T	1,2-Dibromoethane	50.000	54.146	-8.3	95	0.00
62 S	4-Bromofluorobenzene	50.000	51.588	-3.2	93	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	96	0.00
64 T	Tetrachloroethene	50.000	49.210	1.6	90	0.00
65 PM	Chlorobenzene	50.000	48.388	3.2	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.777	-3.6	94	0.00
67 C	Ethyl Benzene	50.000	48.988	2.0#	91	0.00
68 T	m/p-Xylenes	100.000	98.668	1.3	91	0.00
69 T	o-Xylene	50.000	50.181	-0.4	92	0.00
70 T	Styrene	50.000	52.460	-4.9	93	0.00
71 P	Bromoform	50.000	57.289	-14.6	92	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	92	0.00
73 T	Isopropylbenzene	50.000	49.196	1.6	90	0.00
74 T	N-ethyl acetate	50.000	59.235	-18.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.199	-0.4	97	0.00
76 T	1,2,3-Trichloropropane	50.000	52.069	-4.1	97	0.00
77 T	Bromobenzene	50.000	50.727	-1.5	93	0.00
78 T	n-propylbenzene	50.000	50.102	-0.2	91	0.00
79 T	2-Chlorotoluene	50.000	48.826	2.3	91	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.664	-1.3	90	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.845	0.3	94	0.00
82 T	4-Chlorotoluene	50.000	49.585	0.8	91	0.00
83 T	tert-Butylbenzene	50.000	49.350	1.3	89	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.260	-0.5	91	0.00
85 T	sec-Butylbenzene	50.000	50.575	-1.2	91	0.00
86 T	p-Isopropyltoluene	50.000	52.092	-4.2	92	0.00
87 T	1,3-Dichlorobenzene	50.000	50.043	-0.1	92	0.00
88 T	1,4-Dichlorobenzene	50.000	48.633	2.7	91	0.00
89 T	n-Butylbenzene	50.000	51.751	-3.5	91	0.00
90 T	Hexachloroethane	50.000	47.752	4.5	88	0.00
91 T	1,2-Dichlorobenzene	50.000	50.194	-0.4	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	56.240	-12.5	95	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.714	-5.4	91	0.00
94 T	Hexachlorobutadiene	50.000	50.349	-0.7	90	0.00
95 T	Naphthalene	50.000	54.696	-9.4	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
Data File : VX031920.D
Acq On : 06 Oct 2022 15:57
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 07 08:06:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	53.293	-6.6	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6