


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100622\
 Data File : VX031919.D
 Acq On : 06 Oct 2022 15:29
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Title : SW846 8260
 Last Update : Thu Oct 06 14:28:46 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.1	5718	PASS
75	95	30	60	52.9	16704	PASS
95	95	100	100	100.0	31576	PASS
96	95	5	9	7.3	2292	PASS
173	174	0.00	2	0.7	153	PASS
174	95	50	100	73.4	23179	PASS
175	174	5	9	7.6	1759	PASS
176	174	95	101	97.8	22678	PASS
177	176	5	9	6.7	1522	PASS