

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038142.D
 Acq On : 06 Oct 2023 19:37
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 07 04:30:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127797	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221198	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132076	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	104125	50.656	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.320%	
35) Dibromofluoromethane	5.385	113	81063	50.859	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.720%	
50) Toluene-d8	8.647	98	317091	52.063	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.120%	
62) 4-Bromofluorobenzene	11.079	95	128222	55.531	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.060%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	60780	49.272	ug/l	99
3) Chloromethane	1.301	50	59275	47.146	ug/l	97
4) Vinyl Chloride	1.374	62	73017	52.380	ug/l	98
5) Bromomethane	1.599	94	103832	56.830	ug/l	97
6) Chloroethane	1.679	64	63636	61.090	ug/l	97
7) Trichlorofluoromethane	1.886	101	160085	58.140	ug/l	99
8) Diethyl Ether	2.136	74	58827	57.286	ug/l	82
9) 1,1,2-Trichlorotrifluo...	2.325	101	58534	48.696	ug/l	98
10) Methyl Iodide	2.453	142	86141	49.807	ug/l	94
11) Tert butyl alcohol	2.965	59	112092	246.897	ug/l #	93
12) 1,1-Dichloroethene	2.319	96	57578	49.244	ug/l	87
13) Acrolein	2.233	56	79660	272.607	ug/l	99
14) Allyl chloride	2.660	41	99994	47.813	ug/l	95
15) Acrylonitrile	3.062	53	237253	253.630	ug/l	99
16) Acetone	2.380	43	192690	224.613	ug/l	92
17) Carbon Disulfide	2.508	76	145489	43.172	ug/l	100
18) Methyl Acetate	2.703	43	185670	50.469	ug/l	94
19) Methyl tert-butyl Ether	3.117	73	235174	50.502	ug/l	99
20) Methylene Chloride	2.788	84	74355	46.079	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	71405	49.323	ug/l	92
22) Diisopropyl ether	3.764	45	224250	51.161	ug/l	87
23) Vinyl Acetate	3.721	43	832715	259.058	ug/l	97
24) 1,1-Dichloroethane	3.611	63	132828	49.934	ug/l	100
25) 2-Butanone	4.556	43	346763	250.522	ug/l	91
26) 2,2-Dichloropropane	4.471	77	85754	41.291	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	86505	50.172	ug/l	91
28) Bromochloromethane	4.897	49	64827	53.985	ug/l	89
29) Tetrahydrofuran	5.007	42	226769	251.637	ug/l	89
30) Chloroform	5.093	83	152074	52.280	ug/l	97
31) Cyclohexane	5.471	56	109025	48.955	ug/l	94
32) 1,1,1-Trichloroethane	5.385	97	128106	51.417	ug/l	100
36) 1,1-Dichloropropene	5.696	75	107099	48.727	ug/l	98
37) Ethyl Acetate	4.715	43	133018	50.958	ug/l	97
38) Carbon Tetrachloride	5.678	117	108462	50.163	ug/l	97
39) Methylcyclohexane	7.379	83	123192	47.851	ug/l	95
40) Benzene	6.037	78	314749	51.200	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038142.D
 Acq On : 06 Oct 2023 19:37
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 07 04:30:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	66041	50.858	ug/l #	89
42) 1,2-Dichloroethane	6.086	62	126826	51.636	ug/l	100
43) Isopropyl Acetate	6.342	43	198330	50.896	ug/l	96
44) Trichloroethene	7.129	130	81155	50.473	ug/l	99
45) 1,2-Dichloropropane	7.434	63	80980	50.896	ug/l	99
46) Dibromomethane	7.580	93	62395	50.562	ug/l	97
47) Bromodichloromethane	7.824	83	117327	53.445	ug/l	98
48) Methyl methacrylate	7.696	41	97107	52.200	ug/l #	86
49) 1,4-Dioxane	7.659	88	48538	1058.833	ug/l #	89
51) 4-Methyl-2-Pentanone	8.574	43	722717	268.883	ug/l	97
52) Toluene	8.720	92	213684	53.361	ug/l	93
53) t-1,3-Dichloropropene	8.982	75	124457	52.816	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	128105	51.744	ug/l	94
55) 1,1,2-Trichloroethane	9.153	97	91476	53.406	ug/l	94
56) Ethyl methacrylate	9.116	69	140406	53.464	ug/l	93
57) 1,3-Dichloropropane	9.311	76	146817	52.997	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	329674	255.744	ug/l	99
59) 2-Hexanone	9.433	43	583118	269.859	ug/l	95
60) Dibromochloromethane	9.525	129	94443	56.338	ug/l	100
61) 1,2-Dibromoethane	9.610	107	102491	55.333	ug/l	100
64) Tetrachloroethene	9.275	164	72820	48.873	ug/l	95
65) Chlorobenzene	10.079	112	236654	49.137	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	85187	52.306	ug/l	99
67) Ethyl Benzene	10.195	91	430688	50.098	ug/l	99
68) m/p-Xylenes	10.305	106	337344	101.483	ug/l	99
69) o-Xylene	10.640	106	163564	50.415	ug/l	100
70) Styrene	10.659	104	281709	53.319	ug/l	96
71) Bromoform	10.799	173	70797	47.039	ug/l #	100
73) Isopropylbenzene	10.963	105	434776	49.086	ug/l	99
74) N-amyl acetate	10.842	43	190325	50.318	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	165077	50.790	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	175886	58.578	ug/l	90
77) Bromobenzene	11.201	156	105471	48.773	ug/l	93
78) n-propylbenzene	11.305	91	525762	48.545	ug/l	100
79) 2-Chlorotoluene	11.366	91	307550	47.883	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	382390	50.040	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	43446	42.876	ug/l	93
82) 4-Chlorotoluene	11.457	91	371163	48.834	ug/l	98
83) tert-Butylbenzene	11.713	119	374791	50.194	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	384214	50.428	ug/l	99
85) sec-Butylbenzene	11.890	105	482776	49.912	ug/l	99
86) p-Isopropyltoluene	12.012	119	399198	50.023	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	206446	47.531	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	213899	48.237	ug/l	99
89) n-Butylbenzene	12.335	91	381642	48.020	ug/l	97
90) Hexachloroethane	12.542	117	67632	51.115	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	205062	49.116	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	39017	49.651	ug/l	88
93) 1,2,4-Trichlorobenzene	13.591	180	132542	47.332	ug/l	96
94) Hexachlorobutadiene	13.725	225	52474	46.558	ug/l	99
95) Naphthalene	13.774	128	462617	47.439	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	128085	47.240	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
Data File : VX038142.D
Acq On : 06 Oct 2023 19:37
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Oct 07 04:30:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Oct 07 04:30:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 127797
Ion Ratio Lower Upper
168 100
99 59.0 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 49.272 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 85 Resp: 60780
Ion Ratio Lower Upper
85 100
87 32.9 16.7 50.1





#3

Chloromethane

Concen: 47.146 ug/l

RT: 1.301 min Scan# 35

Delta R.T. -0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 50 Resp: 59275

Ion Ratio Lower Upper

50 100

52 33.3 25.5 38.3



#4

Vinyl Chloride

Concen: 52.380 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX038142.D




Acq: 06 Oct 2023 19:37


Tgt Ion: 62 Resp: 73017

Ion Ratio Lower Upper

62 100


64 32.6 25.4 38.0








#5
Bromomethane
Concen: 56.830 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 94 Resp: 103832
Ion Ratio Lower Upper
94 100
96 96.1 74.6 111.8



#6
Chloroethane
Concen: 61.090 ug/l
RT: 1.679 min Scan# 97
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37


Tgt Ion: 64 Resp: 63636
Ion Ratio Lower Upper
64 100
66 32.9 25.1 37.7








#7
 Trichlorofluoromethane
 Concen: 58.140 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX038142.D
 Acq: 06 Oct 2023 19:37

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:101 Resp: 160085
 Ion Ratio Lower Upper
 101 100
 103 63.4 51.4 77.2



#8
 Diethyl Ether
 Concen: 57.286 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.006 min
 Lab File: VX038142.D
 Acq: 06 Oct 2023 19:37

Tgt Ion: 74 Resp: 58827
 Ion Ratio Lower Upper
 74 100
 45 87.6 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.696 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:101	Resp:	58534
Ion	Ratio	Lower Upper
101	100	
85	46.6	37.0 55.6
151	75.2	58.6 87.8



#10
Methyl Iodide
Concen: 49.807 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion:142	Resp:	86141
Ion	Ratio	Lower Upper
142	100	
127	46.6	41.6 62.4
141	14.2	11.6 17.4





#11

Tert butyl alcohol

Concen: 246.897 ug/l

RT: 2.965 min Scan# 307

Delta R.T. 0.006 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 112092

Ion Ratio Lower Upper

59 100

57 10.0 6.0 9.0#



#12

1,1-Dichloroethene

Concen: 49.244 ug/l

RT: 2.319 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37




Tgt Ion: 96 Resp: 57578


Ion Ratio Lower Upper

96 100

61 156.1 144.6 216.8

98 63.2 50.4 75.6





#13

Acrolein

Concen: 272.607 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 79660

Ion Ratio Lower Upper

56 100

55 72.1 57.1 85.7



#14

Allyl chloride

Concen: 47.813 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37




Tgt Ion: 41 Resp: 99994


Ion Ratio Lower Upper

41 100

39 72.0 55.8 83.6


76 35.4 24.3 36.5








#15
Acrylonitrile
Concen: 253.630 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 237253
Ion Ratio Lower Upper
53 100
52 83.6 66.0 99.0
51 36.9 29.8 44.6



#16
Acetone
Concen: 224.613 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 43 Resp: 192690
Ion Ratio Lower Upper
43 100
58 31.4 21.9 32.9









#17
Carbon Disulfide
Concen: 43.172 ug/l
RT: 2.508 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 145489
Ion Ratio Lower Upper
76 100
78 9.0 7.0 10.6



#18
Methyl Acetate
Concen: 50.469 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 43 Resp: 185670
Ion Ratio Lower Upper
43 100
74 25.3 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 50.502 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 235174
Ion Ratio Lower Upper
73 100
57 21.1 17.3 25.9



#20
Methylene Chloride
Concen: 46.079 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37


Tgt Ion: 84 Resp: 74355
Ion Ratio Lower Upper
84 100
49 121.6 103.6 155.4
51 36.9 33.8 50.6
86 66.9 49.4 74.0








#21
trans-1,2-Dichloroethene
Concen: 49.323 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 71405
Ion Ratio Lower Upper
96 100
61 138.9 119.2 178.8
98 65.9 48.7 73.1



#22
Diisopropyl ether
Concen: 51.161 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 45 Resp: 224250
Ion Ratio Lower Upper
45 100
43 72.6 69.7 104.5
87 30.4 20.8 31.2
59 11.8 9.0 13.4









#23
 Vinyl Acetate
 Concen: 259.058 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX038142.D
 Acq: 06 Oct 2023 19:37




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 832715
 Ion Ratio Lower Upper
 43 100
 86 10.9 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 49.934 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX038142.D
 Acq: 06 Oct 2023 19:37

Tgt Ion: 63 Resp: 132828
 Ion Ratio Lower Upper
 63 100
 98 6.3 3.3 9.8
 100 4.3 2.1 6.5









#25
2-Butanone
Concen: 250.522 ug/l
RT: 4.556 min Scan# 569
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 346763
Ion Ratio Lower Upper
43 100
72 28.9 19.4 29.2



#26
2,2-Dichloropropane
Concen: 41.291 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.006 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 77 Resp: 85754
Ion Ratio Lower Upper
77 100
97 23.8 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 50.172 ug/l

RT: 4.489 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 86505

Ion Ratio Lower Upper

96 100

61 139.9 0.0 312.2

98 64.1 0.0 129.0



#28

Bromochloromethane

Concen: 53.985 ug/l

RT: 4.897 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37




Tgt Ion: 49 Resp: 64827


Ion Ratio Lower Upper

49 100

129 2.1 0.0 3.8

130 81.0 57.5 86.3





#29

Tetrahydrofuran

Concen: 251.637 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 226769

Ion Ratio Lower Upper

42 100

72 50.1 34.2 51.4

71 47.0 32.2 48.2



#30

Chloroform

Concen: 52.280 ug/l

RT: 5.093 min Scan# 657

Delta R.T. -0.000 min

Lab File: VX038142.D




Acq: 06 Oct 2023 19:37


Tgt Ion: 83 Resp: 152074

Ion Ratio Lower Upper

83 100


85 65.1 50.3 75.5








#31
Cyclohexane
Concen: 48.955 ug/l
RT: 5.471 min Scan# 719
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 56 Resp: 109025
Ion Ratio Lower Upper
56 100
69 32.7 23.4 35.2
84 88.2 65.9 98.9




#32
1,1,1-Trichloroethane
Concen: 51.417 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37



Tgt Ion: 97 Resp: 128106
Ion Ratio Lower Upper
97 100
99 64.3 51.1 76.7
61 45.4 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 50.656 ug/l

RT: 5.952 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 104125

Ion Ratio Lower Upper

65 100

67 52.0 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37




Tgt Ion: 114 Resp: 221198


Ion Ratio Lower Upper

114 100

63 19.9 0.0 45.0

88 15.6 0.0 36.6









#35
Dibromofluoromethane
Concen: 50.859 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 81063
Ion Ratio Lower Upper
113 100
111 101.3 82.9 124.3
192 18.1 13.3 19.9



#36
1,1-Dichloropropene
Concen: 48.727 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.006 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37


Tgt Ion: 75 Resp: 107099
Ion Ratio Lower Upper
75 100
110 36.1 17.0 50.9
77 30.9 24.7 37.1








#37
Ethyl Acetate
Concen: 50.958 ug/l
RT: 4.715 min Scan# 595
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 133018
Ion Ratio Lower Upper
43 100
61 14.9 10.6 16.0
70 10.9 8.3 12.5



#38
Carbon Tetrachloride
Concen: 50.163 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 117 Resp: 108462
Ion Ratio Lower Upper
117 100
119 94.7 78.2 117.2
121 30.9 24.4 36.6





#39

Methylcyclohexane

Concen: 47.851 ug/l

RT: 7.379 min Scan# 1033

Delta R.T. -0.006 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 123192

Ion Ratio Lower Upper

83 100

55 74.8 64.6 97.0

98 44.9 36.7 55.1



#40

Benzene

Concen: 51.200 ug/l

RT: 6.037 min Scan# 812

Delta R.T. -0.001 min

Lab File: VX038142.D




Acq: 06 Oct 2023 19:37


Tgt Ion: 78 Resp: 314749

Ion Ratio Lower Upper

78 100


77 24.9 19.3 28.9








#41
Methacrylonitrile
Concen: 50.858 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41 Resp: 66041
Ion Ratio Lower Upper
41 100
39 57.9 45.2 67.8
67 81.2 52.6 78.8#
52 34.7 22.7 34.1#



#42
1,2-Dichloroethane
Concen: 51.636 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37


Tgt Ion: 62 Resp: 126826
Ion Ratio Lower Upper
62 100
98 8.0 0.0 16.0








#43
Isopropyl Acetate
Concen: 50.896 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 198330
Ion Ratio Lower Upper
43 100
61 20.2 15.2 22.8
87 14.3 9.8 14.6



#44
Trichloroethene
Concen: 50.473 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion:130 Resp: 81155
Ion Ratio Lower Upper
130 100
95 101.2 0.0 204.4









#45
1,2-Dichloropropane
Concen: 50.896 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 80980
Ion Ratio Lower Upper
63 100
65 29.8 24.3 36.5



#46
Dibromomethane
Concen: 50.562 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 93 Resp: 62395
Ion Ratio Lower Upper
93 100
95 82.9 66.6 100.0
174 91.5 69.4 104.0









#47
Bromodichloromethane
Concen: 53.445 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 117327
Ion Ratio Lower Upper
83 100
85 63.8 49.8 74.6
127 8.6 7.1 10.7



#48
Methyl methacrylate
Concen: 52.200 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37


Tgt Ion: 41 Resp: 97107
Ion Ratio Lower Upper
41 100
69 89.2 59.2 88.8#
39 67.1 47.7 71.5








#49
1,4-Dioxane
Concen: 1058.833 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 88 Resp: 48538
Ion Ratio Lower Upper
88 100
43 31.4 0.0 0.0#
58 65.7 60.1 90.1



#50
Toluene-d8
Concen: 52.063 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.006 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 98 Resp: 317091
Ion Ratio Lower Upper
98 100
100 63.6 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 268.883 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 722717
Ion Ratio Lower Upper
43 100
58 39.8 30.4 45.6



#52
Toluene
Concen: 53.361 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 92 Resp: 213684
Ion Ratio Lower Upper
92 100
91 176.8 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 52.816 ug/l

RT: 8.982 min Scan# 1295

Delta R.T. 0.006 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 124457

Ion Ratio Lower Upper

75 100

77 30.9 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 51.744 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37




Tgt Ion: 75 Resp: 128105


Ion Ratio Lower Upper

75 100

77 31.6 25.3 37.9


39 51.6 46.7 70.1








#55
 1,1,2-Trichloroethane
 Concen: 53.406 ug/l
 RT: 9.153 min Scan# 1323
 Delta R.T. -0.000 min
 Lab File: VX038142.D
 Acq: 06 Oct 2023 19:37


Instrument :
 MSVOA_X
 ClientSampleId :





Tgt Ion: 97	Resp: 91476
Ion Ratio	Lower Upper
97 100	
83 87.2	65.8 98.8
85 56.9	41.6 62.4
99 63.9	47.8 71.8




#56
 Ethyl methacrylate
 Concen: 53.464 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. 0.000 min
 Lab File: VX038142.D
 Acq: 06 Oct 2023 19:37



Tgt Ion: 69	Resp: 140406
Ion Ratio	Lower Upper
69 100	
41 62.6	56.8 85.2
39 40.6	32.0 48.0





#57

1,3-Dichloropropane

Concen: 52.997 ug/l

RT: 9.311 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 146817

Ion Ratio Lower Upper

76 100

78 32.9 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 255.744 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX038142.D




Acq: 06 Oct 2023 19:37


Tgt Ion: 63 Resp: 329674

Ion Ratio Lower Upper

63 100

106 28.7 22.5 33.7





#59

2-Hexanone

Concen: 269.859 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 583118

Ion Ratio Lower Upper

43 100

58 55.6 26.2 78.5



#60

Dibromochloromethane

Concen: 56.338 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX038142.D




Acq: 06 Oct 2023 19:37


Tgt Ion: 129 Resp: 94443

Ion Ratio Lower Upper

129 100

127 77.7 38.8 116.4









#61
1,2-Dibromoethane
Concen: 55.333 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 102491
Ion Ratio Lower Upper
107 100
109 93.9 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 55.531 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 95 Resp: 128222
Ion Ratio Lower Upper
95 100
174 77.4 0.0 135.6
176 74.2 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 230161
Ion Ratio Lower Upper
117 100
82 55.4 46.2 69.4
119 31.8 25.5 38.3



#64
Tetrachloroethene
Concen: 48.873 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion:164 Resp: 72820
Ion Ratio Lower Upper
164 100
166 136.1 102.7 154.1
129 102.9 80.4 120.6
131 100.7 76.5 114.7









#65
Chlorobenzene
Concen: 49.137 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 236654
Ion Ratio Lower Upper
112 100
114 31.6 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 52.306 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion:131 Resp: 85187
Ion Ratio Lower Upper
131 100
133 94.3 47.1 141.4
119 65.7 32.2 96.6





#67

Ethyl Benzene

Concen: 50.098 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 430688

Ion Ratio Lower Upper

91 100

106 31.5 24.9 37.3



#68

m/p-Xylenes

Concen: 101.483 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. 0.000 min

Lab File: VX038142.D




Acq: 06 Oct 2023 19:37


Tgt Ion:106 Resp: 337344

Ion Ratio Lower Upper

106 100

91 204.8 165.5 248.3





#69

o-Xylene

Concen: 50.415 ug/l

RT: 10.640 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:106 Resp: 163564

Ion Ratio Lower Upper

106 100

91 218.4 108.9 326.8



#70

Styrene

Concen: 53.319 ug/l

RT: 10.659 min Scan# 1570

Delta R.T. -0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37




Tgt Ion:104 Resp: 281709


Ion Ratio Lower Upper

104 100

78 50.8 43.6 65.4

103 55.9 43.6 65.4





#71

Bromoform

Concen: 47.039 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:173 Resp: 70797

Ion Ratio Lower Upper

173 100

175 48.0 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37




Tgt Ion:152 Resp: 132076


Ion Ratio Lower Upper

152 100

115 86.4 43.3 129.9

150 172.7 0.0 345.0





#73

Isopropylbenzene

Concen: 49.086 ug/l

RT: 10.963 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 434776

Ion Ratio Lower Upper

105 100

120 26.1 12.9 38.6



#74

N-ethyl acetate

Concen: 50.318 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Tgt Ion: 43 Resp: 190325




Ion Ratio Lower Upper


43 100

70 44.3 32.9 49.3

55 24.9 19.3 28.9

61 24.8 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 50.790 ug/l

RT: 11.213 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 165077

Ion Ratio Lower Upper

83 100

131 9.6 5.3 15.8

85 64.0 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 58.578 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX038142.D




Acq: 06 Oct 2023 19:37


Tgt Ion: 75 Resp: 175886

Ion Ratio Lower Upper

75 100

77 32.3 19.1 57.5





#77

Bromobenzene

Concen: 48.773 ug/l

RT: 11.201 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 105471

Ion Ratio Lower Upper

156 100

77 162.9 88.4 265.3

158 97.8 48.0 144.0



#78

n-propylbenzene

Concen: 48.545 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX038142.D




Acq: 06 Oct 2023 19:37


Tgt Ion: 91 Resp: 525762

Ion Ratio Lower Upper

91 100

120 23.0 11.5 34.4









#79
2-Chlorotoluene
Concen: 47.883 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 307550
Ion Ratio Lower Upper
91 100
126 34.1 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 50.040 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 105 Resp: 382390
Ion Ratio Lower Upper
105 100
120 46.4 24.1 72.3









#81
 trans-1,4-Dichloro-2-butene
 Concen: 42.876 ug/l
 RT: 11.018 min Scan# 1629
 Delta R.T. 0.000 min
 Lab File: VX038142.D
 Acq: 06 Oct 2023 19:37




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 43446
 Ion Ratio Lower Upper
 75 100
 53 102.9 90.5 135.7
 89 46.0 36.9 55.3



#82
 4-Chlorotoluene
 Concen: 48.834 ug/l
 RT: 11.457 min Scan# 1701
 Delta R.T. 0.000 min
 Lab File: VX038142.D
 Acq: 06 Oct 2023 19:37

Tgt Ion: 91 Resp: 371163
 Ion Ratio Lower Upper
 91 100
 126 29.6 14.2 42.6





#83

tert-Butylbenzene

Concen: 50.194 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 374791

Ion Ratio Lower Upper

119 100

91 59.0 31.2 93.6

134 21.8 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 50.428 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX038142.D




Acq: 06 Oct 2023 19:37


Tgt Ion:105 Resp: 384214

Ion Ratio Lower Upper

105 100

120 42.9 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 49.912 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX038142.D
 Acq: 06 Oct 2023 19:37




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 482776
 Ion Ratio Lower Upper
 105 100
 134 19.3 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 50.023 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.006 min
 Lab File: VX038142.D
 Acq: 06 Oct 2023 19:37

Tgt Ion:119 Resp: 399198
 Ion Ratio Lower Upper
 119 100
 134 26.0 12.8 38.3
 91 26.0 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 47.531 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min


Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:146 Resp: 206446

Ion Ratio Lower Upper

146 100

111 42.2 22.1 66.1

148 64.0 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 48.237 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37




Tgt Ion:146 Resp: 213899


Ion Ratio Lower Upper

146 100

111 40.6 21.3 63.7

148 63.7 31.8 95.4





#89

n-Butylbenzene

Concen: 48.020 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 381642

Ion Ratio Lower Upper

91 100

92 51.8 27.4 82.0

134 24.5 12.6 37.8



#90

Hexachloroethane

Concen: 51.115 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.006 min

Lab File: VX038142.D




Acq: 06 Oct 2023 19:37


Tgt Ion: 117 Resp: 67632

Ion Ratio Lower Upper

117 100

201 70.2 32.9 98.6









#91
1,2-Dichlorobenzene
Concen: 49.116 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 205062
Ion Ratio Lower Upper
146 100
111 42.0 22.4 67.0
148 63.6 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.651 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX038142.D
Acq: 06 Oct 2023 19:37

Tgt Ion: 75 Resp: 39017
Ion Ratio Lower Upper
75 100
155 80.3 36.1 108.5
157 102.6 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 47.332 ug/l

RT: 13.591 min Scan# 2050

Delta R.T. 0.006 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 132542

Ion Ratio Lower Upper

180 100

182 93.9 48.3 144.8

145 31.4 18.0 54.0



#94

Hexachlorobutadiene

Concen: 46.558 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37




Tgt Ion:225 Resp: 52474


Ion Ratio Lower Upper

225 100

223 63.4 31.1 93.2

227 64.2 32.5 97.5





#95

Naphthalene

Concen: 47.439 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:128 Resp: 462617

Ion Ratio Lower Upper

128 100

127 13.1 10.3 15.5

129 10.9 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 47.240 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX038142.D

Acq: 06 Oct 2023 19:37

Tgt Ion:180 Resp: 128085

Ion Ratio Lower Upper

180 100

182 94.7 46.9 140.6

145 33.8 17.8 53.5

