

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038117.D
 Acq On : 06 Oct 2023 09:52
 Operator : JC/MD
 Sample : VX1006MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1006MBL01

Quant Time: Oct 07 04:15:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	113728	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198122	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92075	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83226	45.498	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.000%
35) Dibromofluoromethane	5.385	113	63550	44.515	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.040%
50) Toluene-d8	8.647	98	246247	45.140	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.280%#
62) 4-Bromofluorobenzene	11.079	95	89400	43.228	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
Data File : VX038117.D
Acq On : 06 Oct 2023 09:52
Operator : JC/MD
Sample : VX1006MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1006MBL01

Quant Time: Oct 07 04:15:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

