

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038125.D
 Acq On : 06 Oct 2023 13:01
 Operator : JC/MD
 Sample : 04693-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2

Quant Time: Oct 07 04:19:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	194760	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97113	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80310	45.439	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.880%
35) Dibromofluoromethane	5.391	113	60964	43.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.880%
50) Toluene-d8	8.653	98	242320	45.187	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.380%#
62) 4-Bromofluorobenzene	11.079	95	89959	44.249	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
Data File : VX038125.D
Acq On : 06 Oct 2023 13:01
Operator : JC/MD
Sample : 04693-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2

Quant Time: Oct 07 04:19:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

