

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038133.D
 Acq On : 06 Oct 2023 16:08
 Operator : JC/MD
 Sample : 04796-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Oct 07 04:21:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	101911	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	178630	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91863	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73484	44.830	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.660%
35) Dibromofluoromethane	5.391	113	57617	44.764	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.520%
50) Toluene-d8	8.647	98	226420	46.034	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.060%
62) 4-Bromofluorobenzene	11.079	95	84659	45.402	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
Data File : VX038133.D
Acq On : 06 Oct 2023 16:08
Operator : JC/MD
Sample : 04796-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Oct 07 04:21:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

