

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038135.D
 Acq On : 06 Oct 2023 16:54
 Operator : JC/MD
 Sample : 04794-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-BASIN-C-20231005

Quant Time: Oct 07 04:22:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	100196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180761	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74171	46.024	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.040%
35) Dibromofluoromethane	5.385	113	56752	43.572	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.140%
50) Toluene-d8	8.653	98	228094	45.828	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.660%#
62) 4-Bromofluorobenzene	11.079	95	83667	44.341	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.680%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1843	1.956	ug/l	94
16) Acetone	2.379	43	1483	2.205	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
Data File : VX038135.D
Acq On : 06 Oct 2023 16:54
Operator : JC/MD
Sample : 04794-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-BASIN-C-20231005

Quant Time: Oct 07 04:22:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

