

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038138.D
 Acq On : 06 Oct 2023 18:04
 Operator : JC/MD
 Sample : 04794-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-SP-303-20231005

Quant Time: Oct 07 04:23:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	94892	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	168097	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69703	45.669	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.340%
35) Dibromofluoromethane	5.385	113	53768	44.391	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.780%
50) Toluene-d8	8.653	98	212665	45.947	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.900%#
62) 4-Bromofluorobenzene	11.079	95	79551	45.336	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.680%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	3621	4.057	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
Data File : VX038138.D
Acq On : 06 Oct 2023 18:04
Operator : JC/MD
Sample : 04794-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-SP-303-20231005

Quant Time: Oct 07 04:23:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

