

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038156.D
 Acq On : 07 Oct 2023 04:06
 Operator : JC/MD
 Sample : 04802-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 231005066-01

Quant Time: Oct 07 06:37:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X100623W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Oct 07 06:30:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	9703	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	73345	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	89050	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	41369	31.208	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.033%
60) 4-Bromofluorobenzene	11.079	95	54098	29.163	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	97.200%
63) Toluene-d8	8.652	98	98630	28.602	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	95.333%
Target Compounds						Qvalue
15) Acetone	2.379	58	2210	9.691	ug/l	96
16) Carbon Disulfide	2.513	76	1368	0.477	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
Data File : VX038156.D
Acq On : 07 Oct 2023 04:06
Operator : JC/MD
Sample : 04802-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
231005066-01

Quant Time: Oct 07 06:37:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X100623W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Oct 07 06:30:32 2023
Response via : Initial Calibration

