

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038116.D
 Acq On : 06 Oct 2023 09:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 07 04:14:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.483	0.503	-4.1	96	0.00
3 P	Chloromethane	0.492	0.473	3.9	98	0.00
4 C	Vinyl Chloride	0.545	0.550	-0.9#	95	0.00
5 T	Bromomethane	0.744	0.832	-11.8	102	0.00
6 T	Chloroethane	0.484	0.432	10.7	93	0.00
7 T	Trichlorofluoromethane	1.077	1.085	-0.7	95	0.00
8 T	Diethyl Ether	0.402	0.400	0.5	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.470	0.476	-1.3	96	0.00
10 T	Methyl Iodide	0.677	0.671	0.9	94	0.00
11 T	Tert butyl alcohol	0.196	0.162	17.3	87	0.00
12 CM	1,1-Dichloroethene	0.457	0.456	0.2#	94	0.00
13 T	Acrolein	0.114	0.116	-1.8	107	0.00
14 T	Allyl chloride	0.818	0.821	-0.4	95	0.00
15 T	Acrylonitrile	0.366	0.361	1.4	94	0.00
16 T	Acetone	0.336	0.309	8.0	91	0.00
17 T	Carbon Disulfide	1.318	1.234	6.4	93	0.00
18 T	Methyl Acetate	1.439	1.442	-0.2	96	0.00
19 T	Methyl tert-butyl Ether	1.822	1.823	-0.1	94	0.00
20 T	Methylene Chloride	0.631	0.584	7.4	96	0.00
21 T	trans-1,2-Dichloroethene	0.566	0.561	0.9	92	0.00
22 T	Diisopropyl ether	1.715	1.771	-3.3	95	0.00
23 T	Vinyl Acetate	1.258	1.331	-5.8	96	0.00
24 P	1,1-Dichloroethane	1.041	1.036	0.5	95	0.00
25 T	2-Butanone	0.542	0.538	0.7	95	0.00
26 T	2,2-Dichloropropane	0.813	0.856	-5.3	97	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.676	-0.1	96	0.00
28 T	Bromochloromethane	0.470	0.496	-5.5	95	0.00
29 T	Tetrahydrofuran	0.353	0.352	0.3	96	0.00
30 C	Chloroform	1.138	1.177	-3.4#	95	0.00
31 T	Cyclohexane	0.871	0.910	-4.5	96	0.00
32 T	1,1,1-Trichloroethane	0.975	1.013	-3.9	94	0.00
33 S	1,2-Dichloroethane-d4	0.804	0.769	4.4	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.360	0.341	5.3	87	0.00
36 T	1,1-Dichloropropene	0.497	0.499	-0.4	96	0.00
37 T	Ethyl Acetate	0.590	0.610	-3.4	97	0.00
38 T	Carbon Tetrachloride	0.489	0.512	-4.7	95	0.00
39 T	Methylcyclohexane	0.582	0.624	-7.2	98	0.00
40 TM	Benzene	1.390	1.441	-3.7	95	0.00
41 T	Methacrylonitrile	0.294	0.313	-6.5	98	0.00
42 TM	1,2-Dichloroethane	0.555	0.572	-3.1	95	0.00
43 T	Isopropyl Acetate	0.881	0.922	-4.7	94	0.00
44 TM	Trichloroethene	0.363	0.384	-5.8	97	0.00
45 C	1,2-Dichloropropane	0.360	0.377	-4.7#	97	0.00
46 T	Dibromomethane	0.279	0.285	-2.2	96	0.00
47 T	Bromodichloromethane	0.496	0.541	-9.1	95	0.00
48 T	Methyl methacrylate	0.420	0.446	-6.2	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038116.D
 Acq On : 06 Oct 2023 09:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 07 04:14:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.011	-10.0	93	0.00
50 S	Toluene-d8	1.377	1.331	3.3	89	0.00
51 T	4-Methyl-2-Pentanone	0.608	0.642	-5.6	97	0.00
52 CM	Toluene	0.905	0.958	-5.9#	96	0.00
53 T	t-1,3-Dichloropropene	0.533	0.599	-12.4	97	0.00
54 T	cis-1,3-Dichloropropene	0.560	0.613	-9.5	96	0.00
55 T	1,1,2-Trichloroethane	0.387	0.405	-4.7	96	0.00
56 T	Ethyl methacrylate	0.594	0.641	-7.9	96	0.00
57 T	1,3-Dichloropropane	0.626	0.669	-6.9	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.291	0.302	-3.8	94	0.00
59 T	2-Hexanone	0.488	0.526	-7.8	99	0.00
60 T	Dibromochloromethane	0.379	0.423	-11.6	98	0.00
61 T	1,2-Dibromoethane	0.419	0.447	-6.7	96	0.00
62 S	4-Bromofluorobenzene	0.522	0.531	-1.7	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.324	0.330	-1.9	96	0.00
65 PM	Chlorobenzene	1.046	1.052	-0.6	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.374	-5.6	95	0.00
67 C	Ethyl Benzene	1.868	1.948	-4.3#	98	0.00
68 T	m/p-Xylenes	0.722	0.757	-4.8	99	0.00
69 T	o-Xylene	0.705	0.731	-3.7	99	0.00
70 T	Styrene	1.148	1.245	-8.4	100	0.00
71 P	Bromoform	0.292	0.322	-10.3	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.353	3.408	-1.6	100	0.00
74 T	N-amyl acetate	1.432	1.528	-6.7	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.230	1.274	-3.6	102	0.00
76 T	1,2,3-Trichloropropane	1.137	1.027	9.7	89	0.00
77 T	Bromobenzene	0.819	0.830	-1.3	101	0.00
78 T	n-propylbenzene	4.100	4.236	-3.3	102	0.00
79 T	2-Chlorotoluene	2.432	2.426	0.2	101	0.00
80 T	1,3,5-Trimethylbenzene	2.893	3.021	-4.4	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.356	0.373	-4.8	99	0.00
82 T	4-Chlorotoluene	2.877	2.946	-2.4	102	0.00
83 T	tert-Butylbenzene	2.827	2.909	-2.9	101	0.00
84 T	1,2,4-Trimethylbenzene	2.884	3.032	-5.1	101	0.00
85 T	sec-Butylbenzene	3.662	3.866	-5.6	102	0.00
86 T	p-Isopropyltoluene	3.021	3.228	-6.9	104	0.00
87 T	1,3-Dichlorobenzene	1.644	1.682	-2.3	103	0.00
88 T	1,4-Dichlorobenzene	1.679	1.715	-2.1	105	0.00
89 T	n-Butylbenzene	3.009	3.220	-7.0	104	0.00
90 T	Hexachloroethane	0.501	0.546	-9.0	102	0.00
91 T	1,2-Dichlorobenzene	1.581	1.653	-4.6	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.297	0.319	-7.4	101	0.00
93 T	1,2,4-Trichlorobenzene	1.060	1.105	-4.2	106	0.00
94 T	Hexachlorobutadiene	0.427	0.450	-5.4	107	0.00
95 T	Naphthalene	3.692	3.770	-2.1	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
Data File : VX038116.D
Acq On : 06 Oct 2023 09:24
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 07 04:14:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.026	1.064	-3.7	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6