

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
 Data File : VX038153.D
 Acq On : 07 Oct 2023 02:56
 Operator : JC/MD
 Sample : VX1006WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1006WBL02

Quant Time: Oct 07 06:36:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X100623W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Oct 07 06:30:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	9603	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	74117	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	95561	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	41651	31.748	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.833%
60) 4-Bromofluorobenzene	11.079	95	55349	27.804	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.667%
63) Toluene-d8	8.653	98	111374	30.098	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.333%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100623\
Data File : VX038153.D
Acq On : 07 Oct 2023 02:56
Operator : JC/MD
Sample : VX1006WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1006WBL02

Quant Time: Oct 07 06:36:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X100623W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Oct 07 06:30:32 2023
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 4.897 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX038153.D
Acq: 07 Oct 2023 02:56




Instrument :
MSVOA_X
ClientSampleId :
VX1006WBL02


Tgt Ion:128 Resp: 9603
Ion Ratio Lower Upper
128 100
49 156.9 0.0 406.3
130 127.0 0.0 328.0



#27
1,2-Dichloroethane-d4
Concen: 31.748 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX038153.D
Acq: 07 Oct 2023 02:56

Tgt Ion: 65 Resp: 41651
Ion Ratio Lower Upper
65 100
67 51.8 42.1 63.1





#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 6.763 min Scan# 91

Delta R.T. 0.000 min

Lab File: VX038153.D

Acq: 07 Oct 2023 02:56

Instrument :

MSVOA_X

ClientSampleId :

VX1006WBL02





Tgt Ion:114 Resp: 74117

Ion Ratio Lower Upper

114 100

63 20.9 16.6 24.8

88 15.7 12.9 19.3



#57

Chlorobenzene-d5

Concen: 30.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX038153.D

Acq: 07 Oct 2023 02:56




Tgt Ion:117 Resp: 95561


Ion Ratio Lower Upper

117 100

82 57.3 44.7 67.1

119 32.0 25.6 38.4









#60
4-Bromofluorobenzene
Concen: 27.804 ug/l
RT: 11.079 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX038153.D
Acq: 07 Oct 2023 02:56

Instrument :
MSVOA_X
ClientSampleId :
VX1006WBL02

Tgt Ion: 95 Resp: 55349
Ion Ratio Lower Upper
95 100
174 80.0 60.5 90.7
176 76.8 58.2 87.4



#63
Toluene-d8
Concen: 30.098 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. 0.006 min
Lab File: VX038153.D
Acq: 07 Oct 2023 02:56

Tgt Ion: 98 Resp: 111374
Ion Ratio Lower Upper
98 100
100 62.4 51.6 77.4

