

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
 Data File : VX048010.D
 Acq On : 06 Oct 2025 09:23
 Operator : JC/MD
 Sample : VX1006MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1006MBL01

Quant Time: Oct 07 02:06:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.532	168	107152	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.739	114	213083	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	218675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	106799	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	92384	54.165	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.340%
35) Dibromofluoromethane	5.367	113	77844	54.473	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.940%
50) Toluene-d8	8.629	98	263593	53.307	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.620%
62) 4-Bromofluorobenzene	11.061	95	109639	58.423	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
Data File : VX048010.D
Acq On : 06 Oct 2025 09:23
Operator : JC/MD
Sample : VX1006MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1006MBL01

Quant Time: Oct 07 02:06:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

