

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
 Data File : VX048011.D
 Acq On : 06 Oct 2025 09:44
 Operator : JC/MD
 Sample : VX1006WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1006WBL01

Quant Time: Oct 07 02:06:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	163752	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	333598	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	343343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	171990	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	135132	51.844	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.680%
35) Dibromofluoromethane	5.373	113	110438	49.362	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.720%
50) Toluene-d8	8.635	98	378416	48.882	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.760%
62) 4-Bromofluorobenzene	11.061	95	161318	54.907	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
Data File : VX048011.D
Acq On : 06 Oct 2025 09:44
Operator : JC/MD
Sample : VX1006WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1006WBL01

Quant Time: Oct 07 02:06:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

