

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
 Data File : VX048018.D
 Acq On : 06 Oct 2025 13:04
 Operator : JC/MD
 Sample : Q3279-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 07 02:12:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	121633	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	244927	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	262334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	129379	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	102930	53.164	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.320%
35) Dibromofluoromethane	5.373	113	81556	49.650	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.300%
50) Toluene-d8	8.634	98	277085	48.750	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.500%
62) 4-Bromofluorobenzene	11.067	95	122105	56.606	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
Data File : VX048018.D
Acq On : 06 Oct 2025 13:04
Operator : JC/MD
Sample : Q3279-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 07 02:12:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

