

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
 Data File : VX048021.D
 Acq On : 06 Oct 2025 14:07
 Operator : JC/MD
 Sample : Q3201-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW15

Quant Time: Oct 07 02:13:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	134028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	273148	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	290347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	149616	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	114515	53.678	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	107.360%	
35) Dibromofluoromethane	5.373	113	91472	49.933	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.860%	
50) Toluene-d8	8.635	98	313165	49.405	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.820%	
62) 4-Bromofluorobenzene	11.067	95	139428	57.959	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	115.920%	
Target Compounds						
16) Acetone	2.386	43	2563	2.763	ug/l	Qvalue # 68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
Data File : VX048021.D
Acq On : 06 Oct 2025 14:07
Operator : JC/MD
Sample : Q3201-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW15

Quant Time: Oct 07 02:13:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

