

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
 Data File : VX048024.D
 Acq On : 06 Oct 2025 15:10
 Operator : JC/MD
 Sample : Q3263-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 266380

Quant Time: Oct 07 02:15:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	166031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	333294	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	344096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	172196	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	138052	52.237	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.480%	
35) Dibromofluoromethane	5.379	113	112981	50.545	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.100%	
50) Toluene-d8	8.635	98	383784	49.620	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.240%	
62) 4-Bromofluorobenzene	11.061	95	164789	56.139	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.280%	
Target Compounds						
					Qvalue	
30) Chloroform	5.093	83	4749	1.121	ug/l	94
47) Bromodichloromethane	7.812	83	11514	3.018	ug/l	99
60) Dibromochloromethane	9.506	129	21403	7.882	ug/l	99
71) Bromoform	10.787	173	13756	6.829	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
Data File : VX048024.D
Acq On : 06 Oct 2025 15:10
Operator : JC/MD
Sample : Q3263-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
266380

Quant Time: Oct 07 02:15:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

