

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
 Data File : VX048026.D
 Acq On : 06 Oct 2025 15:51
 Operator : JC/MD
 Sample : Q3202-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-2D

Quant Time: Oct 07 02:16:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	132008	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	274950	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	286333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	145340	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	114478	54.481	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.960%	
35) Dibromofluoromethane	5.373	113	92366	50.091	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.180%	
50) Toluene-d8	8.634	98	316591	49.619	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.240%	
62) 4-Bromofluorobenzene	11.067	95	137249	56.679	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.360%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.307	50	737	0.410	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
Data File : VX048026.D
Acq On : 06 Oct 2025 15:51
Operator : JC/MD
Sample : Q3202-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-2D

Quant Time: Oct 07 02:16:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

