

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
 Data File : VX048027.D
 Acq On : 06 Oct 2025 16:12
 Operator : JC/MD
 Sample : Q3202-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-5

Quant Time: Oct 07 02:17:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	119710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	246037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	256820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	131834	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	100626	52.809	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.620%
35) Dibromofluoromethane	5.379	113	82014	49.704	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.400%
50) Toluene-d8	8.635	98	272664	47.756	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.520%
62) 4-Bromofluorobenzene	11.067	95	121370	56.012	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.020%
Target Compounds						
16) Acetone	2.380	43	1213	1.464	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
Data File : VX048027.D
Acq On : 06 Oct 2025 16:12
Operator : JC/MD
Sample : Q3202-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-5

Quant Time: Oct 07 02:17:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

