

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
 Data File : VX048031.D
 Acq On : 06 Oct 2025 17:35
 Operator : JC/MD
 Sample : Q3202-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-3D

Quant Time: Oct 07 02:20:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	124280	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	255750	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	273710	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	139753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	112472	56.855	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.720%	
35) Dibromofluoromethane	5.373	113	90381	52.694	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.380%	
50) Toluene-d8	8.635	98	297038	50.049	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.100%	
62) 4-Bromofluorobenzene	11.067	95	129057	57.297	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 114.600%	
Target Compounds						
3) Chloromethane	1.307	50	635	0.375	ug/l #	88
16) Acetone	2.386	43	1695	1.971	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
Data File : VX048031.D
Acq On : 06 Oct 2025 17:35
Operator : JC/MD
Sample : Q3202-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-3D

Quant Time: Oct 07 02:20:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

