

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100719\
 Data File : VX012830.D
 Acq On : 07 Oct 2019 11:57
 Operator : JC/SP
 Sample : VX1007WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1007WBL01

Quant Time: Oct 08 11:35:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	175278	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	298497	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	257116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	104935	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	113585	44.50	ug/l	0.00
Spiked Amount			Recovery	=	89.00%	
35) Dibromofluoromethane	5.48	113	89621	50.41	ug/l	0.00
Spiked Amount			Recovery	=	100.82%	
50) Toluene-d8	8.71	98	350631	51.22	ug/l	0.00
Spiked Amount			Recovery	=	102.44%	
62) 4-Bromofluorobenzene	11.13	95	125287	46.54	ug/l	0.00
Spiked Amount			Recovery	=	93.08%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100719\
Data File : VX012830.D
Acq On : 07 Oct 2019 11:57
Operator : JC/SP
Sample : VX1007WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1007WBL01

Quant Time: Oct 08 11:35:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

