

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100719\
 Data File : VX012842.D
 Acq On : 07 Oct 2019 16:47
 Operator : JC/SP
 Sample : K5154-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-105S

Quant Time: Oct 08 12:13:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178950	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	303011	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	260802	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	102339	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	118548	45.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.98%	
35) Dibromofluoromethane	5.49	113	92996	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
50) Toluene-d8	8.71	98	358627	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
62) 4-Bromofluorobenzene	11.13	95	125507	45.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.86%	
Target Compounds						
16) Acetone	2.44	43	7572	5.486	ug/l #	89
65) Chlorobenzene	10.14	112	47890	8.720	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100719\
Data File : VX012842.D
Acq On : 07 Oct 2019 16:47
Operator : JC/SP
Sample : K5154-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-105S

Quant Time: Oct 08 12:13:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

