

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100719\
 Data File : VX012847.D
 Acq On : 07 Oct 2019 18:44
 Operator : JC/SP
 Sample : K5189-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ157I-20191001

Quant Time: Oct 08 12:22:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	179442	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	311282	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	257936	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	97515	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119069	45.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.14%	
35) Dibromofluoromethane	5.48	113	92826	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
50) Toluene-d8	8.71	98	366177	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
62) 4-Bromofluorobenzene	11.13	95	122676	43.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.40%	
Target Compounds						
16) Acetone	2.43	43	5835	4.216	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100719\
Data File : VX012847.D
Acq On : 07 Oct 2019 18:44
Operator : JC/SP
Sample : K5189-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ157I-20191001

Quant Time: Oct 08 12:22:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

