

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100719\
 Data File : VX012848.D
 Acq On : 07 Oct 2019 19:07
 Operator : JC/SP
 Sample : K5190-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FC-MW09S-20191002

Quant Time: Oct 08 12:23:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178729	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	305057	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	258550	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	99978	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	119008	45.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.46%	
35) Dibromofluoromethane	5.49	113	91171	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
50) Toluene-d8	8.71	98	362323	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
62) 4-Bromofluorobenzene	11.14	95	121199	44.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.12%	
Target Compounds						
16) Acetone	2.43	43	6766	4.908	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100719\
Data File : VX012848.D
Acq On : 07 Oct 2019 19:07
Operator : JC/SP
Sample : K5190-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FC-MW09S-20191002

Quant Time: Oct 08 12:23:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

