

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100719\
 Data File : VX012849.D
 Acq On : 07 Oct 2019 19:30
 Operator : JC/SP
 Sample : K5190-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FC-MW09I-20191002

Quant Time: Oct 08 12:25:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	179325	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	307885	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	264373	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	105920	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	120814	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
35) Dibromofluoromethane	5.49	113	92026	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
50) Toluene-d8	8.71	98	361760	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
62) 4-Bromofluorobenzene	11.14	95	122309	44.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.10%	
Target Compounds						
16) Acetone	2.43	43	8959	6.477	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100719\
Data File : VX012849.D
Acq On : 07 Oct 2019 19:30
Operator : JC/SP
Sample : K5190-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FC-MW09I-20191002

Quant Time: Oct 08 12:25:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

