

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100719\
 Data File : VX012861.D
 Acq On : 08 Oct 2019 01:19
 Operator : JC/SP
 Sample : K5232-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-GMZ1CSS-TB1

Quant Time: Oct 08 13:00:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	180333	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	308638	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	262300	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	105021	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	118584	45.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.32%	
35) Dibromofluoromethane	5.49	113	90752	49.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.74%	
50) Toluene-d8	8.71	98	362963	51.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.56%	
62) 4-Bromofluorobenzene	11.13	95	127655	45.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100719\
Data File : VX012861.D
Acq On : 08 Oct 2019 01:19
Operator : JC/SP
Sample : K5232-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-GMZ1CSS-TB1

Quant Time: Oct 08 13:00:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

