

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100719\
 Data File : VX012868.D
 Acq On : 08 Oct 2019 04:03
 Operator : JC/SP
 Sample : K5216-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0541

Quant Time: Oct 08 13:12:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	180448	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	309774	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	263439	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	109278	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	119054	45.31	ug/l	0.00
Spiked Amount			Recovery	=	90.62%	
35) Dibromofluoromethane	5.49	113	92001	49.87	ug/l	0.00
Spiked Amount			Recovery	=	99.74%	
50) Toluene-d8	8.71	98	361702	50.91	ug/l	0.00
Spiked Amount			Recovery	=	101.82%	
62) 4-Bromofluorobenzene	11.14	95	130259	46.63	ug/l	0.00
Spiked Amount			Recovery	=	93.26%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	549932	237.352	ug/l	95
12) 1,1-Dichloroethene	2.36	96	7847	4.045	ug/l	97
16) Acetone	2.43	43	2082	1.496	ug/l	94
21) trans-1,2-Dichloroethene	3.16	96	163767	76.192	ug/l	97
27) cis-1,2-Dichloroethene	4.59	96	3230439	1295.293	ug/l	82
44) Trichloroethene	7.23	130	15379913	6847.909	ug/l	92
52) Toluene	8.78	92	38222	6.922	ug/l	100
67) Ethyl Benzene	10.25	91	5842	0.581	ug/l	94
68) m/p-Xylenes	10.36	106	5208	1.382	ug/l	85
69) o-Xylene	10.70	106	2409	0.661	ug/l	85
78) n-propylbenzene	11.35	91	2411	0.241	ug/l	98
80) 1,3,5-Trimethylbenzene	11.50	105	3218	0.431	ug/l	99
84) 1,2,4-Trimethylbenzene	11.81	105	6062	0.805	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100719\
Data File : VX012868.D
Acq On : 08 Oct 2019 04:03
Operator : JC/SP
Sample : K5216-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0541

Quant Time: Oct 08 13:12:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

