

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100719\
 Data File : VX012873.D
 Acq On : 08 Oct 2019 05:59
 Operator : JC/SP
 Sample : K5122-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-MMW3414

Quant Time: Oct 08 13:20:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	175678	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	301665	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	261969	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	102622	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	117815	46.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.10%	
35) Dibromofluoromethane	5.49	113	89364	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
50) Toluene-d8	8.71	98	358188	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
62) 4-Bromofluorobenzene	11.14	95	123825	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
Target Compounds						
16) Acetone	2.44	43	6000	4.428	ug/l	89
37) Ethyl Acetate	4.83	43	5136	1.511	ug/l #	90
44) Trichloroethene	7.21	130	4219	1.929	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100719\
Data File : VX012873.D
Acq On : 08 Oct 2019 05:59
Operator : JC/SP
Sample : K5122-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-MMW3414

Quant Time: Oct 08 13:20:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

