

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100719\
 Data File : VX012875.D
 Acq On : 08 Oct 2019 06:46
 Operator : JC/SP
 Sample : K5122-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-CHY07

Quant Time: Oct 08 13:22:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	179274	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	301979	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	265825	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	107922	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	118921	45.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.10%	
35) Dibromofluoromethane	5.49	113	92022	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
50) Toluene-d8	8.71	98	361958	52.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.52%	
62) 4-Bromofluorobenzene	11.14	95	127735	46.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.82%	
Target Compounds						
16) Acetone	2.44	43	9754	7.054	ug/l	94
44) Trichloroethene	7.21	130	3207	1.465	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100719\
Data File : VX012875.D
Acq On : 08 Oct 2019 06:46
Operator : JC/SP
Sample : K5122-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-CHY07

Quant Time: Oct 08 13:22:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

