


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX100719\
Data File : VX012855.D
Acq On : 07 Oct 2019 21:50
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Title : SW846 8260
Last Update : Mon Sep 23 12:27:01 2019



AutoFind: Scans 1680, 1681, 1682; Background Corrected with Scan 1673

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.6	7443	PASS
75	95	30	60	47.8	21493	PASS
95	95	100	100	100.0	44954	PASS
96	95	5	9	6.4	2859	PASS
173	174	0.00	2	0.7	221	PASS
174	95	50	100	73.2	32917	PASS
175	174	5	9	7.6	2494	PASS
176	174	95	101	96.5	31776	PASS
177	176	5	9	5.7	1814	PASS