

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100721\
 Data File : VX024644.D
 Acq On : 07 Oct 2021 14:32
 Operator : JC/MD
 Sample : M4001-02 100X
 Misc : 5.01g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EW917

Quant Time: Oct 07 23:07:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML100721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 07 23:05:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	277947	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	222421	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	111002	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	83330	48.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.84%
7) Chloroethane-d5	1.666	69	54003	47.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.34%
11) 1,1-Dichloroethene-d2	2.306	63	121214	35.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.74%
21) 2-Butanone-d5	4.465	46	140574	99.36	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.36%
24) Chloroform-d	5.068	84	164798	47.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.56%
26) 1,2-Dichloroethane-d4	5.964	65	120226	48.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.54%
32) Benzene-d6	5.983	84	357287	57.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.04%
36) 1,2-Dichloropropane-d6	7.318	67	109307	55.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	111.56%
41) Toluene-d8	8.653	98	330564	54.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.30%
43) trans-1,3-Dichloroprop...	8.952	79	48626	50.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.16%
47) 2-Hexanone-d5	9.391	63	101751	110.60	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.60%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	138767	51.42	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.84%
66) 1,2-Dichlorobenzene-d4	12.323	152	112676	51.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.76%
Target Compounds						
					Qvalue	
25) Chloroform	5.099	83	11554	3.02	ug/L	98
34) Trichloroethene	7.135	95	20762	10.97	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100721\
Data File : VX024644.D
Acq On : 07 Oct 2021 14:32
Operator : JC/MD
Sample : M4001-02 100X
Misc : 5.01g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EW917

Quant Time: Oct 07 23:07:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML100721WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 07 23:05:28 2021
Response via : Initial Calibration

