

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100722\  
 Data File : VX031951.D  
 Acq On : 07 Oct 2022 10:45  
 Operator : JC/MD  
 Sample : VX1007WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX1007WBL01

Quant Time: Oct 08 05:07:16 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X100622W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 06 14:28:46 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 63164    | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 98894    | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 105732   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 65471    | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 65163    | 49.954   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =     | 99.900%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 54841    | 48.990   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 97.980%  |
| 50) Toluene-d8              | 8.647  | 98    | 196853   | 48.870   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =     | 97.740%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 68516    | 48.114   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 96.220%  |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX100722\  
Data File : VX031951.D  
Acq On : 07 Oct 2022 10:45  
Operator : JC/MD  
Sample : VX1007WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1007WBL01

Quant Time: Oct 08 05:07:16 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X100622W.M  
Quant Title : SW846 8260  
QLast Update : Thu Oct 06 14:28:46 2022  
Response via : Initial Calibration

