

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031958.D
 Acq On : 07 Oct 2022 13:38
 Operator : JC/MD
 Sample : N4883-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB092622

Quant Time: Oct 08 05:10:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	57282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	92391	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58129	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63403	53.595	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.200%
35) Dibromofluoromethane	5.385	113	52321	50.029	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.060%
50) Toluene-d8	8.653	98	186979	49.686	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.380%
62) 4-Bromofluorobenzene	11.079	95	63417	47.668	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031958.D
Acq On : 07 Oct 2022 13:38
Operator : JC/MD
Sample : N4883-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB092622

Quant Time: Oct 08 05:10:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

