

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031959.D
 Acq On : 07 Oct 2022 14:02
 Operator : JC/MD
 Sample : N4883-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB092722

Quant Time: Oct 08 05:10:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	59377	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	97572	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	103468	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62297	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66780	54.458	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.920%
35) Dibromofluoromethane	5.385	113	54868	49.679	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	8.647	98	198056	49.835	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.680%
62) 4-Bromofluorobenzene	11.079	95	66547	47.364	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031959.D
Acq On : 07 Oct 2022 14:02
Operator : JC/MD
Sample : N4883-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB092722

Quant Time: Oct 08 05:10:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

