

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031961.D
 Acq On : 07 Oct 2022 14:48
 Operator : JC/MD
 Sample : N5050-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-01-160-180-100622

Quant Time: Oct 08 05:11:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	53592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	87270	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	92332	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57051	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58846	53.168	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.340%
35) Dibromofluoromethane	5.385	113	47715	48.302	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.600%
50) Toluene-d8	8.647	98	175342	49.328	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.660%
62) 4-Bromofluorobenzene	11.079	95	60321	48.001	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.000%
Target Compounds						
52) Toluene	8.720	92	12135	4.839	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031961.D
Acq On : 07 Oct 2022 14:48
Operator : JC/MD
Sample : N5050-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-01-160-180-100622

Quant Time: Oct 08 05:11:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

