

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031962.D
 Acq On : 07 Oct 2022 15:11
 Operator : JC/MD
 Sample : N5050-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-01-184-204-100622

Quant Time: Oct 08 05:12:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	57015	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	90343	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	95882	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61338	52.093	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.180%
35) Dibromofluoromethane	5.385	113	51339	50.203	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.400%
50) Toluene-d8	8.647	98	181528	49.331	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.660%
62) 4-Bromofluorobenzene	11.079	95	62815	48.286	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.580%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1035	2.586	ug/l	94
52) Toluene	8.720	92	29248	11.265	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031962.D
Acq On : 07 Oct 2022 15:11
Operator : JC/MD
Sample : N5050-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-01-184-204-100622

Quant Time: Oct 08 05:12:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

