

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031963.D
 Acq On : 07 Oct 2022 15:35
 Operator : JC/MD
 Sample : N5050-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-01-135-155-100622

Quant Time: Oct 08 05:12:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	54696	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	87472	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	93850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56834	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60689	53.727	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.460%	
35) Dibromofluoromethane	5.391	113	49993	50.491	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.980%	
50) Toluene-d8	8.653	98	179296	50.324	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.640%	
62) 4-Bromofluorobenzene	11.079	95	60833	48.297	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.600%	
Target Compounds						
52) Toluene	8.720	92	10481	4.169	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031963.D
Acq On : 07 Oct 2022 15:35
Operator : JC/MD
Sample : N5050-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-01-135-155-100622

Quant Time: Oct 08 05:12:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

