

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031965.D
 Acq On : 07 Oct 2022 16:21
 Operator : JC/MD
 Sample : N5058-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-808-810

Quant Time: Oct 08 06:06:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	56239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	89499	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96125	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58198	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60636	52.207	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.420%	
35) Dibromofluoromethane	5.391	113	51065	50.406	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.820%	
50) Toluene-d8	8.653	98	182709	50.120	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.240%	
62) 4-Bromofluorobenzene	11.079	95	61479	47.704	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.400%	
Target Compounds						
44) Trichloroethene	7.129	130	65914	57.146	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031965.D
Acq On : 07 Oct 2022 16:21
Operator : JC/MD
Sample : N5058-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-808-810

Quant Time: Oct 08 06:06:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

