

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031966.D
 Acq On : 07 Oct 2022 16:45
 Operator : JC/MD
 Sample : N4883-06DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-138DL

Quant Time: Oct 08 05:13:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	53414	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	86876	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	92430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59518	53.955	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.900%
35) Dibromofluoromethane	5.385	113	50124	50.971	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.940%
50) Toluene-d8	8.647	98	172582	48.772	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.540%
62) 4-Bromofluorobenzene	11.079	95	59026	47.184	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.360%
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	3.111	73	88019	32.239	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031966.D
Acq On : 07 Oct 2022 16:45
Operator : JC/MD
Sample : N4883-06DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-138DL

Quant Time: Oct 08 05:13:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

