

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031967.D
 Acq On : 07 Oct 2022 17:08
 Operator : JC/MD
 Sample : N4882-20 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-132

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/10/2022
 Supervised By : Mahesh Dadoda 10/10/2022

Quant Time: Oct 08 05:14:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	57699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	94052	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60763	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63248	53.078	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.160%			
35) Dibromofluoromethane	5.391	113	53453	50.209	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.420%			
50) Toluene-d8	8.653	98	187889	49.046	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.100%			
62) 4-Bromofluorobenzene	11.079	95	64010	47.264	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.520%			
Target Compounds						
11) Tert butyl alcohol	2.983	59	6751m	45.536	ug/l	
19) Methyl tert-butyl Ether	3.111	73	93492	31.700	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031967.D
Acq On : 07 Oct 2022 17:08
Operator : JC/MD
Sample : N4882-20 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-132

Quant Time: Oct 08 05:14:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/10/2022
Supervised By : Mahesh Dadoda 10/10/2022

