

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031980.D
 Acq On : 08 Oct 2022 00:00
 Operator : JC/MD
 Sample : N5061-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-100722

Quant Time: Oct 08 06:32:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	52913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	83728	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	89886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57526	52.643	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.280%	
35) Dibromofluoromethane	5.385	113	48399	51.067	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.140%	
50) Toluene-d8	8.647	98	167377	49.079	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.160%	
62) 4-Bromofluorobenzene	11.079	95	58007	48.113	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.220%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031980.D
Acq On : 08 Oct 2022 00:00
Operator : JC/MD
Sample : N5061-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-100722

Quant Time: Oct 08 06:32:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

