

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031982.D
 Acq On : 08 Oct 2022 00:46
 Operator : JC/MD
 Sample : N5061-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-01-204-224-100722

Quant Time: Oct 08 06:33:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	51266	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	81915	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	90780	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57088	53.920	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.840%			
35) Dibromofluoromethane	5.385	113	47484	51.211	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.420%			
50) Toluene-d8	8.647	98	168618	50.537	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.080%			
62) 4-Bromofluorobenzene	11.079	95	57285	48.565	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.140%			
Target Compounds						
52) Toluene	8.720	92	18193	7.728	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031982.D
Acq On : 08 Oct 2022 00:46
Operator : JC/MD
Sample : N5061-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-01-204-224-100722

Quant Time: Oct 08 06:33:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

