

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031983.D
 Acq On : 08 Oct 2022 01:09
 Operator : JC/MD
 Sample : N5061-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-01-204-224-100722-FD

Quant Time: Oct 08 06:34:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	52802	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	85291	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	90803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56882	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59494	54.558	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.120%			
35) Dibromofluoromethane	5.385	113	48506	50.242	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.480%			
50) Toluene-d8	8.652	98	170327	49.029	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.060%			
62) 4-Bromofluorobenzene	11.079	95	58672	47.772	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.540%			
Target Compounds						
16) Acetone	2.391	43	1183	3.192	ug/l #	86
52) Toluene	8.720	92	17521	7.148	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031983.D
Acq On : 08 Oct 2022 01:09
Operator : JC/MD
Sample : N5061-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-01-204-224-100722-FD

Quant Time: Oct 08 06:34:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

