

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031987.D
 Acq On : 08 Oct 2022 02:42
 Operator : JC/MD
 Sample : N4937-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-117

Quant Time: Oct 08 06:35:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	51933	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	81587	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	87472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53592	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55825	52.050	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.100%
35) Dibromofluoromethane	5.385	113	47288	51.204	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.400%
50) Toluene-d8	8.647	98	166019	49.958	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.079	95	56252	47.881	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.760%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	826	2.266	ug/l	95
30) Chloroform	5.087	83	3381	2.001	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031987.D
Acq On : 08 Oct 2022 02:42
Operator : JC/MD
Sample : N4937-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-117

Quant Time: Oct 08 06:35:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

