

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031989.D
 Acq On : 08 Oct 2022 03:28
 Operator : JC/MD
 Sample : N4937-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-123

Quant Time: Oct 08 06:36:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	52949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	83868	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	91386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59556	54.463	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.920%	
35) Dibromofluoromethane	5.391	113	48708	51.307	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.620%	
50) Toluene-d8	8.647	98	172764	50.574	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.140%	
62) 4-Bromofluorobenzene	11.079	95	58799	48.688	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.380%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.977	59	3824	28.107	ug/l #	73
16) Acetone	2.386	43	2133	5.739	ug/l	96
19) Methyl tert-butyl Ether	3.111	73	98772	36.495	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031989.D
Acq On : 08 Oct 2022 03:28
Operator : JC/MD
Sample : N4937-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-123

Quant Time: Oct 08 06:36:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

