

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031992.D
 Acq On : 08 Oct 2022 04:38
 Operator : JC/MD
 Sample : N4937-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-093022

Quant Time: Oct 08 06:37:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	52801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	83771	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	91113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55246	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58520	53.666	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.340%
35) Dibromofluoromethane	5.391	113	48799	51.463	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.920%
50) Toluene-d8	8.647	98	168162	49.284	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.560%
62) 4-Bromofluorobenzene	11.079	95	57860	47.966	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031992.D
Acq On : 08 Oct 2022 04:38
Operator : JC/MD
Sample : N4937-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-093022

Quant Time: Oct 08 06:37:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

