

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
 Data File : VX031951.D
 Acq On : 07 Oct 2022 10:45
 Operator : JC/MD
 Sample : VX1007WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1007WBL01

Quant Time: Oct 08 05:07:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	63164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	98894	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65471	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65163	49.954	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.900%
35) Dibromofluoromethane	5.385	113	54841	48.990	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.980%
50) Toluene-d8	8.647	98	196853	48.870	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.740%
62) 4-Bromofluorobenzene	11.079	95	68516	48.114	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.220%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100722\
Data File : VX031951.D
Acq On : 07 Oct 2022 10:45
Operator : JC/MD
Sample : VX1007WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1007WBL01

Quant Time: Oct 08 05:07:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 711
Delta R.T. -0.000 min
Lab File: VX031951.D
Acq: 07 Oct 2022 10:45




Instrument :
MSVOA_X
ClientSampleId :
VX1007WBL01


Tgt Ion:168 Resp: 63164
Ion Ratio Lower Upper
168 100
99 56.8 48.8 73.2



#33
1,2-Dichloroethane-d4
Concen: 49.954 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.006 min
Lab File: VX031951.D
Acq: 07 Oct 2022 10:45

Tgt Ion: 65 Resp: 65163
Ion Ratio Lower Upper
65 100
67 52.3 0.0 102.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. -0.000 min
Lab File: VX031951.D
Acq: 07 Oct 2022 10:45




Instrument :
MSVOA_X
ClientSampleId :
VX1007WBL01


Tgt Ion:114 Resp: 98894
Ion Ratio Lower Upper
114 100
63 17.5 0.0 42.6
88 15.1 0.0 33.0



#35
Dibromofluoromethane
Concen: 48.990 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX031951.D
Acq: 07 Oct 2022 10:45

Tgt Ion:113 Resp: 54841
Ion Ratio Lower Upper
113 100
111 103.4 82.4 123.6
192 18.6 14.1 21.1









#50
Toluene-d8
Concen: 48.870 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX031951.D
Acq: 07 Oct 2022 10:45




Instrument :
MSVOA_X
ClientSampleId :
VX1007WBL01


Tgt Ion: 98 Resp: 196853
Ion Ratio Lower Upper
98 100
100 67.6 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 48.114 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX031951.D
Acq: 07 Oct 2022 10:45

Tgt Ion: 95 Resp: 68516
Ion Ratio Lower Upper
95 100
174 81.6 0.0 153.2
176 78.6 0.0 144.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX031951.D
Acq: 07 Oct 2022 10:45

Instrument :
MSVOA_X
ClientSampleId :
VX1007WBL01

Tgt Ion:117 Resp: 105732
Ion Ratio Lower Upper
117 100
82 50.7 46.8 70.2
119 32.2 26.4 39.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX031951.D
Acq: 07 Oct 2022 10:45

Tgt Ion:152 Resp: 65471
Ion Ratio Lower Upper
152 100
115 60.8 43.4 130.1
150 156.1 0.0 346.8

